

Manual Supplement

Manual Title: 1507/1503 Users
Part Number: 2416013
Print Date: June 2005
Revision/Date:

Supplement Issue: **8**
Issue Date: 9/16
Page Count: 3

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On page 4, under **Battery Saver (*Sleep Mode*)**, replace the last sentence in the first paragraph with the following:

The Tester comes out of Sleep mode when the rotary switch is turned to off and back on.

On page 24, under Model 1503, replace the Display Range, Output Voltage 500V:

From: 0.1 to 20.00 M Ω

To: 0.01 to 20.00 M Ω

Change #2



On page 5, under Table 2, add the following:

Warning

To avoid shock, injury, or damage to the Tester, use only accessories with 1000V CAT III or 600V CAT IV rating.

Change #3, 134

On page 3, add the following to the **Symbols** table:

	Conforms to relevant South Korean EMC Standards.
CAT II	Measurement Category II is applicable to test and measuring circuits directly to utilization points (socket outlets and similar points) low-voltage MAINS installation.
CAT III	Measurement Category III is applicable to test and measuring circuits connected to the distribution part of the building's low-voltage MAINS installation.
CAT IV	Measurement Category IV is applicable to test and measuring circuits connected at the source of the building's low-voltage MAINS installation.
 660V	WARNING. Do not apply greater than 660 Volts

Change #4, 61173

On page 21, under **General Specifications**, replace **Altitude** with:

Altitude..... Operating: 2000 m CAT IV 600 V, Non Operating (Storage): 12,000 m

Change #5, 64919, 65176, 367, 539

On page 20, under **General Specifications**, remove **Shock, Electromagnetic Compatibility, Safety, Certifications**, and add:

Electromagnetic Compatibility (EMC)

International IEC 61326-1: Portable Electromagnetic Environment
 CISPR 11: Group 1, Class A

Group 1: Equipment has intentionally generated and/or uses conductively-coupled radio frequency energy that is necessary for the internal function of the equipment itself.

Class A: Equipment is suitable for use in all establishments other than domestic and those directly connected to a low-voltage power supply network that supplies buildings used for domestic purposes. There may be potential difficulties in ensuring electromagnetic compatibility in other environments due to conducted and radiated disturbances.

Emissions that exceed the levels required by CISPR 11 can occur when the equipment is connected to a test object.

Korea (KCC) Class A Equipment (Industrial Broadcasting & Communication Equipment)

Class A: Equipment meets requirements for industrial electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and not to be used in homes.

USA (FCC)..... 47 CFR 15 subpart B. This product is considered an exempt device per clause 15.103.

Safety

General IEC 61010-1: Pollution Degree 2

Measurement..... IEC 61010-2-030; CAT IV 600 V

Compliance to EN 61557 Parts 1 and 2