

Supplement

Title:	373 Users	Supplement Issue:	5
Part Number:	3612805	Issue Date:	3/15
Print Date:	July 2010	Page Count:	3
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above Instruction Sheet.

Change #1, 55883

On page 15, under *DC Voltage*, add flag note [1] to Accuracy and add:

- [1] In an RF field greater than 1 V/m between 130 MHz and 350 MHz, add 65 counts to specified accuracy.



Change #2, 56248, 64117

On page 3, under **Warning**,

Measure a known voltage first to make sure that the Meter operates correctly.

Change #3

On page 6, add the following to the **Symbols** table:

	Conforms to CAN/CSA-C22.2 No. 61010-1, second edition, including Amendment 1.
	Conforms to relevant South Korean EMC Standards.

Change #4, 64117, 64124, 488

On page 17, delete EMI, EMC, Safety Compliance, and the Agency Approvals sections and add the following new sections:

Electromagnetic Compatibility (EMC)

International..... IEC 61326-1: Portable Electromagnetic Environment

CISPR 11: Group 1, Class A

Group 1: Equipment has intentionally generated and/or uses conductively-coupled radio frequency energy that is necessary for the internal function of the equipment itself.

Class A: Equipment is suitable for use in all establishments other than domestic and those directly connected to a low-voltage power supply network that supplies buildings used for domestic purposes. There may be potential difficulties in ensuring electromagnetic compatibility in other environments due to conducted and radiated disturbances.

Emissions that exceed the levels required by CISPR 11 can occur when the equipment is connected to a test object.

Korea (KCC) Class A Equipment (Industrial Broadcasting & Communication Equipment)

Class A: Equipment meets requirements for industrial electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and not to be used in homes.

USA (FCC) 47 CFR 15 subpart B. This product is considered an exempt device per clause 15.103.

Safety..... IEC 61010-1: 300 V CAT IV / 600 V CAT III, Pollution Degree 2