

# Supplement

Manual Title: 721Ex Safety Sheet  
Part Number: 4561164 Supplement Issue: 1  
Print Date: March 2016 Issue Date: 6/16  
Revision/Date: Page Count: 1

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1, 560

Under **Safety Specifications**, following Ex-Hazardous Area Markings add:

Entity Parameter ..... Measurement Jacks:

- $U_i = 30 \text{ V}$
- $I_i = 80 \text{ mA}$
- $P_i = 750 \text{ mW}$
- $C_i = 0 \text{ }\mu\text{F}$
- $L_i = 0 \text{ mH}$
- $U_o = 7.14 \text{ V}$
- $I_o = 1.12 \text{ mA}$
- $P_o = 2 \text{ mW}$
- $C_o = 240 \text{ }\mu\text{F}$
- $L_o = 1 \text{ H}$

Circular Connector: For use with LTP100A RTD Probe only.