

# Manual Supplement

Manual Title:	805/805 FC Users	Supplement Issue:	1
Print Date:	May 2012	Issue Date:	12/18
Revision/Date:	2, 12/14	Page Count:	1

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.  
This manual is distributed as an electronic manual on the following CD-ROM:

CD Title: 805  
CD Rev. & Date: 1, 9/2012  
CD PN: 4045063

## Change #1, 638

On page 4, under **Infrared Thermometer**, replace the entire Accuracy (typical) specification with:

Accuracy (typical)..... $\pm 5$  °C from -20 °C to 50 °C,  $\pm 10$  °C from 50 °C to 150 °C