

FLUKE®

80T-IR

Manual Supplement

This supplement applies to:

80T-IR Instruction Sheet

PN 935072, Feb 1993, Rev.3, 9/97

80T-IR Instruction Card

PN 935077, April 1993, Rev.2,11/96

To ensure accuracy of your instruction sheet, make the changes as described in this supplement.

Change #1

Under the Specifications section, replace the Ambient Operating Range with:

Ambient Operating Range: 0 to 50°C (32 to 122°F)