

Supplement

Title: T5-600/T5-1000 Safety Sheet

Part Number: 1631798

Supplement Issue: 1

Print Date: December 2001

Issue Date: 8/18

Revision/Date: 2, 8/18

Page Count: 2

Change #1, 613















Under **Warnings**, remove the 14th bullet and replace the 20th and the 21st bullets with:

Warning

- Do not use test leads if they are damaged. Examine the test leads for damaged insulation.
- Measure a known voltage first to make sure that the Product operates correctly.

Replace the **Symbols** table with:

Symbols

	Consult user documentation.
	Hazardous voltage. Risk of electric shock.
	Risk of Danger. Important information. See Manual.
	Double Insulated
	Battery
	Conforms to relevant North American Safety Standards.
	Conforms to EU directives.
	Earth
	DC (Direct Current)
	AC (Alternating Current)
	Conforms to relevant South Korean EMC Standards.
	Conforms to relevant Australian EMC requirements.
	Certified by TÜV SÜD Product Service.
CAT II	Measurement Category II is applicable to test and measuring circuits connected directly to utilization points (socket outlets and similar points) of the low-voltage MAINS installation.
CAT III	Measurement Category III is applicable to test and measuring circuits connected to the distribution part of the building's low-voltage MAINS installation.
CAT IV	Measurement Category IV is applicable to test and measuring circuits connected at the source of the building's low-voltage MAINS installation.
	This product complies with the WEEE Directive (2002/96/EC) marking requirements. The affixed label indicates that you must not discard this electrical/electronic product in domestic household waste. Product Category: With reference to the equipment types in the WEEE Directive Annex I, this product is classed as category 9 "Monitoring and Control Instrumentation" product. Do not dispose of this product as unsorted municipal waste. Go to Fluke's website for recycling information.

Add **Safety Specifications** section:

Safety Specifications

Safety	IEC 61010-1, IEC 61010-2-032, IEC 61010-2-033: Pollution Degree 2
Maximum Voltage Between any Terminal and Earth Ground	
T5-600	CAT III, 600 V rms
T5-1000	CAT III, 1000 V rms; CAT IV, 600 V rms,
Maximum Voltage Between Current Fork and Earth Ground	
T5-600 and T5-1000	CAT III, 1000 V rms; CAT IV, 600 V rms
Operating Temperature	-10 °C to +50 °C
Operating Altitude	2000 m
Maximum Relative Humidity	95 % (5 °C to 30 °C); 75 % (30 °C to 40 °C); 45 % (40 °C to 50 °C)