

Supplement


Manual Title:	i800 Instruction Sheet		
Part Number:	2842289	Supplement Issue:	1
Print Date:	June 2007	Issue Date:	5/19
Revision/Date:	1, 8/08	Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 622

Under **Introduction**, remove the 3rd paragraph.

Under **Symbols**, replace the last two rows in the table with:

	Conforms to relevant Australian Safety and EMC standards.
CAT III	Measurement Category III is applicable to test and measuring circuits connected to the distribution part of the building's low-voltage MAINS installation.

On the last page, remove the **Safety Standards** and the **EMC Standards** sections, and replace with:

Safety

General IEC 61010-1: Pollution Degree 2

Measurement IEC 61010-2-032: CAT III 600 V

Electromagnetic Environment

IEC 61326-1..... Portable

CISPR 11: Group 1, Class A

Group 1: Equipment has intentionally generated and/or use conductively coupled radio-frequency energy which is necessary for the internal functioning of the equipment itself.

Class A: Equipment is suitable for use in all establishments other than domestic and those directly connected to a low voltage power supply network which supplies buildings used for domestic purposes. There may be potential difficulties in ensuring electromagnetic compatibility in other environments, due to conducted and radiated disturbances.

Caution: This equipment is not intended for use in residential environments and may not provide adequate protection to radio reception in such environments.